



# SIVA INSTRUMENTS

## Million Megohm Meter

Model : LS3B / LS-3C



### SPECIFICATIONS

Test Voltage	: LS-3B = 63V / 100V / 500V selectable LS-3C = 100V / 250V/ 500V / 1000V Selectable Voltages can be modified as per customer request
Accuracy	: $\pm 0.5\% \pm 1V$
Stability	: $\pm 0.01\%$
Resistance Range	: 10e6 Ohm to 10e15 Ohms in seven ranges
Accuracy	: $\pm 2\%$ up to 10e9 and $\pm 5\%$ above 10e9

### SIVANANDA ELECTRONICS

DEEPAK MAHAL, LAM ROAD, DEVLALI-422 401, NASHIK. MAHARASHTRA, INDIA.

Phone : 0253-2491423, 2491504, 2491816 E-mail : mkb@sivanandaelectronics.com Fax: 0253-2492291

**MUMBAI** : 208, MIDAS, Sahar Plaza Complex, Andheri-Kurla Road, Andheri (E), MUMBAI-400 059,

☐PHONE:022-56430590 / 91 / 92 FAX: 022-28389232.

**NEW DELHI** : G -145, Kalkaji, New Delhi -110 019. ☐PHONE : 6221582, 6434442

MFRS. OF : ELECTRONIC INSTRUMENTS, METAL DETECTORS, TURNSTILES, COMPUTERISED ATTENDANCE RECORDING & ACCESS CONTROL  
VISIT US AT [www.sivanandaelctro.com](http://www.sivanandaelctro.com)